

# MANAGING PATIENTS WITH ICDs

## ICDs and electromagnetic interference


### Key take-away points

- Most electrical household products should be kept at least 15 cm from the ICD.
- Sources of EMI significant enough to cause issues are uncommon.
- Cooking on induction cook tops and arc welding should be avoided.



Electricity → EMI

### 15 cm rule



> 15 cm



### Notes